

Endura-Seals Injection Port Septa



Description	Item No.
Endura-Seals Injection Port Septa Kit, 11mm (7/16")	E11-Kit
Endura-Seals Injection Port Septa Replacement, 11mm (7/16")	SE11
Endura-Seals Injection Port Septa Kit, Shimadzu type	SESH-Kit
Endura-Seals Injection Port Septa Replacement, Shimadzu type	SESH



Warnings

Warnings must be read carefully and understood. Improper use of this product can cause harm or death to personnel and damage to property!



Wear eye protection and use caution when working with pressurized systems.

When using a high-pressure gas source overpressure protection must be provided.

Injection Port and the Endura-Seal nut may be hot after use.

Max Pressure: 100 PSI. Max Temp 400C.

For use with 23 gauge cone tip needles only.

Special precautions might be required when using hydrogen or hazardous gases. Consult local regulations and your company's safety procedures.

Setup and Use

Endura-Seals Installation

1. Start with a clean injection port, and inlet pressure at zero and turned off.
2. Gently press the Endura-Seal into the injection port.
3. Screw on the included nut to finger tight plus $\frac{1}{4}$ turn. Do not overtighten.